

Compare list of calibrator function

|                         | GPS-925            | FLUKE 726          |
|-------------------------|--------------------|--------------------|
| Accuracy                | 0.01               | 0.01               |
| Current source          | Y                  | Y                  |
| Simulate transmitter    | Y                  | Y                  |
| Frequency               | Y                  | Y                  |
| Output Range            | 0-5V adjustable    | 5V, not adjustable |
| Pulse count             | Y                  | N                  |
| Pressure Input          | N                  | External           |
| Cold Junction           | Man/Auto           | External           |
| Resolution              | $5\frac{1}{2}$ bit | $4\frac{1}{2}$ bit |
| Numeric Keyboard        | Y                  | N                  |
| Base Stability          | 3PPM               | 15PPM              |
| Cold Junction Stability | 0.2° C             | 0.2° C             |
| Auto-Calibration        | Y                  | N                  |

An obvious advantage compared with FLUKE:

Internal Lithium battery with super capacity, also supply with ac adapter